



Ref. Certif. No.

DK-130807-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Open frame Medical power supply unit (PSU)

Name and address of the applicant

SELEC CONTROLS PVT LTD
EL 27/1, ELECTRONIC ZONE TTC
INDUSTRIAL AREA, MIDC
MAHAPE NAVI MUMBAI, Maharashtra 400709
India

Name and address of the manufacturer

SELEC CONTROLS PVT LTD
EL 27/1, ELECTRONIC ZONE TTC
INDUSTRIAL AREA, MIDC
MAHAPE NAVI MUMBAI, Maharashtra 400709
India

Name and address of the factory

Selec Controls Pvt Ltd
EL30, Electronic Zone
TTC Industrial Area
MIDC Mahape Navi Mumbai, Maharashtra 400710
India

Note: When more than one factory, please report on page 2

☐ Additional Information on page 2

Ratings and principal characteristics

Input: 100-240 V ~, 47- 63 Hz, 0.85 A to 0.42 A Output: See CB Test Report for rating details.

Trademark (if any)

SELEC

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

OPS2x3-40-XXX-VCYY-Z-CU, Class II

☒ Additional Information on page 2

Additional information (if necessary may also be reported on page 2)

Additionally evaluated to: EN 60601-1:2006, EN 60601-1:2006/A1:2013, EN 60601-1:2006/A12:2014

National Differences specified in the CB Test Report.

The risk management requirements of the standard were not addressed

☐ Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012

As shown in the Test Report Ref. No. which forms part of this Certificate

E525579-D1002-1/A0/C0-CB issued on 2022-08-05

This CB Test Certificate is issued by the National Certification Body



- ☐ UL (US), 333 Plingsten Rd IL 60062, Northbrook, USA
- ☒ UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- ☐ UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- ☐ UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2022-08-08

Signature:

Jan-Erik Storgaard



Ref. Certif. No.

DK-130807-UL

Additional Model Detail(s):

OPS 40 series: OPS2x3-40-XXX-VCYY-Z-CU, As per model Nomenclature - OPS2x4-40-XXX-VCYY-Z-CU

OPS- Represent Series name (Open frame power supply)

2x3 -Represent dimension in inches (2width x 3 Length).

40- Represent the maximum output power of family in watts.

XXX- Represent the typical output voltage in VDC .(First two XX represents the typical output voltage in VDC (05 V DC to 48 V DC) and last X represent fractional value of output voltage, can be blank or any number from 1 to 9.).

V- Represent the input voltage range (Blank for 100-240 V AC).

C- Represent the connector type (Any alphabet indicating type/ make of output connector).

YY- Represent the minor output voltage Variation (Blank – Standard product or Any alphanumeric combination indicating minor output voltage variation in specification).

Z-Represent Classification of the device (2 for Class II (without earth)).

CU- Represent the Selec internal code (UL, cUI approved).

Additional information (if necessary)



☐ UL (US), 333 Plingsten Rd IL 60062, Northbrook, USA

☒ UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

☐ UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

☐ UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2022-08-08

Signature:

Jan-Erik Storgaard