



Ref. Certif. No.

DK-130568-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Open frame medical power supply unit (PSU)

Name and address of the applicant

SELEC CONTROLS PVT LTD  
EL 27/1, ELECTRONIC ZONE TTC  
INDUSTRIAL AREA, MIDC  
MAHAPE NAVI MUMBAI, Maharashtra 400709  
INDIA

Name and address of the manufacturer

SELEC CONTROLS PVT LTD  
EL 27/1, ELECTRONIC ZONE TTC  
INDUSTRIAL AREA, MIDC  
MAHAPE NAVI MUMBAI, Maharashtra 400709  
INDIA

Name and address of the factory

SELEC CONTROLS PVT LTD  
EL 30, ELECTRONIC ZONE , TTC INDUSTRIAL  
AREA, MIDC MAHAPE,  
NAVI MUMBIA, MAHARASTTRA- 400710  
INDIA

Note: When more than one factory, please report on page 2

Additional Information on page 2

Ratings and principal characteristics

Input: 100-240 V ~, 47- 63 Hz, 1.4 A to 0.8 An  
Output: see test report for details.

Trademark (if any)



Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

OPS2x4-60-XXX-VCYY-Z-CU  
 Additional Information on page 2

Additional information (if necessary may also be reported on page 2)

**Additionally evaluated to:** EN 60601-1:2006, EN 60601-1:2006/A1:2013, EN 60601-1:2006/A12:2014  
National Differences specified in the CB Test Report.  
Class I

A sample of the product was tested and found to be in conformity with

The risk management requirements of the standard were not addressed  
 Additional Information on page 2

As shown in the Test Report Ref. No. which forms part of this Certificate

IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012

E525579-D1001-1/A0/C0-CB issued on 2022-07-28

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2022-08-03

Signature:

Jan-Erik Storgaard



Ref. Certif. No.

**DK-130568-UL**

**Additional Model Detail(s):**

**OPS 60 Series:** OPS2x4-60-XXX-VCYY-Z-CU,

As per model Nomenclature - OPS2x4-60-XXX-VCYY-Z-CU

OPS- Represent Series name (Open frame power supply)

2x4 -Represent dimension in inches ( 2width x 4 Length).

60- Represent the maximum output power of family in watts.

XXX- Represent the typical output voltage in VDC .( First two XX represents the typical output voltage in VDC (05 VDC to 48V DC) and last X represent fractional value of output voltage, can be blank or any number from 1 to 9.).

V- Represent the input voltage range ( Blank for 100-240 V AC ).

C- Represent the connector type ( Any alphabet indicating type/ make of output connector).

YY- Represent the minor output voltage Variation ( Blank – Standard product or Any alphanumeric combination indicating minor output voltage variation in specification).

Z-Represent Classification of the device (1 for Class I (with earth) )

CU- Represent the Selec internal code ( UL, cUL approved).

**Additional information (if necessary)**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2022-08-03

Signature:

Jan-Erik Storgaard