

Ref. Certif. No.

DK-130595-A1-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME **CB TEST CERTIFICATE** Product Open frame Medical power supply unit (PSU) SELEC CONTROLS PVT LTD Name and address of the applicant EL 27/1, ELECTRONIC ZONE TTC INDUSTRIAL AREA, MIDC MAHAPE NAVI MUMBAI, Maharashtra 400709 India SELEC CONTROLS PVT LTD Name and address of the manufacturer EL 27/1, ELECTRONIC ZONE TTC INDUSTRIAL AREA, MIDC MAHAPE NAVI MUMBAI, Maharashtra 400709 India Selec Controls Pvt Ltd Name and address of the factory EL30, Electronic Zone **TTC Industrial Area** MIDC Mahape Navi Mumbai, Maharashtra 400710 Note: When more than one factory, please report on page 2 India Additional Information on page 2 Input: 100-240 V ~, 47- 63 Hz, 0.85 A to 0.42 A Output: See CB Test Report Ratings and principal characteristics for rating details. Trademark (if any) SELEC Customer's Testing Facility (CTF) Stage used OPS2x3-40-XXX-VCYY-Z-CU, Class I Model / Type Ref. Additional Information on page 2 Additionally evaluated to: EN 60601-1:2006, EN 60601-1:2006/A1:2013, Additional information (if necessary may also be EN 60601-1:2006/A12:2014 reported on page 2) National Differences specified in the CB Test Report. The certificate was revised to include corrections. The risk management requirements of the standard were not addressed Additional Information on page 2 A sample of the product was tested and found IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 to be in conformity with As shown in the Test Report Ref. No. which forms E525579-D1002-1/A0/C0-CB issued on 2022-08-05 part of this Certificate This CB Test Certificate is issued by the National Certification Body UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA For full legal entity names see www.ul.com/ncbnames Date: 2022-08-08 Signature: Original Issue Date: 2022-08-05 Jan-Erik Storgaard





Additional Model Detail(s):

OPS 40 series: OPS2x3-40-XXX-VCYY-Z-CU, As per model Nomenclature - OPS2x4-40-XXX-VCYY-Z-CU

OPS- Represent Series name (Open frame power supply)

2x3 -Represent dimension in inches (2width x 3 Length).

40- Represent the maximum output power of family in watts.

XXX- Represent the typical output voltage in VDC .(First two XX represents the typical output voltage in VDC (05 V DC to 48 V DC) and last X represent fractional value of output voltage, can be blank or any number from 1 to 9.).

V- Represent the input voltage range (Blank for 100-240 V AC).

C- Represent the connector type (Any alphabet indicating type/ make of output connector).

YY- Represent the minor output voltage Variation (Blank – Standard product or Any alphanumeric combination indicating minor output voltage variation in specification).

Z-Represent Classification of the device (1 for Class 1 (with earth)).

CU- Represent the Selec internal code (UL, cUI approved).

Summary of Modifications: This is correction to certificate only to add Class I information that was previously omitted.

Additional information (if necessary)



□ UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA 図 UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK □ UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN □ UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2022-08-08 Original Issue Date: 2022-08-05

Jan out Superiod Signature:

Jan-Erik Storgaard