

DK-130636-UL

### IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

## **CB TEST CERTIFICATE**

**Product** 

Name and address of the applicant

Name and address of the manufacturer

Name and address of the factory

Note: When more than one factory, please report on page 2

Ratings and principal characteristics

Trademark (if any)

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

Additional information (if necessary may also be reported on page 2)

A sample of the product was tested and found to be in conformity with

As shown in the Test Report Ref. No. which forms part of this Certificate

Open frame medical power supply unit (PSU)

SELEC CONTROLS PVT LTD EL 27/1, ELECTRONIC ZONE TTC INDUSTRIAL AREA, MIDC MAHAPE NAVI MUMBAI, Maharashtra 400709 **INDIA** 

SELEC CONTROLS PVT LTD EL 27/1, ELECTRONIC ZONE TTC INDUSTRIAL AREA, MIDC MAHAPE NAVI MUMBAI, Maharashtra 400709 **INDIA** 

SELEC CONTROLS PVT LTD EL 30, ELECTRONIC ZONE, TTC INDUSTRIAL AREA, MIDC MAHAPE, NAVI MUMBIA, MAHARASTTRA- 400710 **INDIA** 

☐ Additional Information on page 2

Input: 100-240 V ~, 47-63 Hz, 1.4 A to 0.8 An Output: see test report for details.



OPS2x4-60-XXX-VCYY-Z-CU □ Additional Information on page 2

Additionally evaluated to: EN 60601-1:2006, EN 60601-1:2006/A1:2013, EN 60601-1:2006/A12:2014 National Differences specified in the CB Test Report.

Class II

The risk management requirements of the standard were not addressed ☐ Additional Information on page 2

IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012

E525579-D1001-1/A0/C0-CB issued on 2022-07-28

This CB Test Certificate is issued by the National Certification Body



Date: 2022-08-03

□ UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

☑ UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
□ UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
□ UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Signature:

Jan-Erik Storgaard



# **DK-130636-UL**

#### Additional Model Detail(s):

OPS 60 Series: OPS2x4-60-XXX-VCYY-Z-CU,

As per model Nomenclature - OPS2x4-60-XXX-VCYY-Z-CU

OPS- Represent Series name (Open frame power supply)

2x4 -Represent dimension in inches (2width x 4 Length).

60- Represent the maximum output power of family in watts.

XXX- Represent the typical output voltage in VDC .( First two XX represents the typical output voltage in VDC (05 VDC to 48V DC) and last X represent fractional value of output voltage, can be blank or any number from 1 to 9.).

V- Represent the input voltage range (Blank for 100-240 V AC).

C- Represent the connector type (Any alphabet indicating type/ make of output connector).

YY- Represent the minor output voltage Variation (Blank - Standard product or Any alphanumeric combination indicating minor output voltage variation in specification).

Z-Represent Classification of the device (2 for Class II (without earth)).

CU- Represent the Selec internal code (UL, cUL approved).

## Additional information (if necessary)



□ UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

■ UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

□ UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN □ UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2022-08-03

Signature: Jan-Erik Storgaard