

# CERTIFICATE OF COMPLIANCE

Certificate Number **2022 07 27 E525579**  
Report Reference **E525579 D1000 1/A0/CO UL**  
Date **2022 07 27**

Issued to: **SELEC CONTROLS PVT LTD**  
Applicant Company: **EL 27/1, ELECTRONIC ZONE TIC INDUSTRIAL AREA, MDC  
MAHAPE  
NAVIMUMBAI, Maharashtra, 400709 INDIA**

Listed Company: **Same as Applicant**

This is to certify that  
representative samples of

**Open frame medical power supply unit (PSU)  
OPS3-5 350 XXXVCYY-Z-CU, OPS2-4 200 XXXVCYY-Z-CU  
(See General Product information for series variants explanation)**

**Have been investigated by UL in accordance with the component  
requirements in the Standard(s) indicated on this Certificate. UL  
Recognized components are incomplete in certain constructional  
features or restricted in performance capabilities and are  
intended for installation in complete equipment submitted for  
investigation to UL LLC.**

Standard(s) for Safety: **ANSI/AAMIES 60601-1:2005(R)2012 and A1:2012,  
C1:2009(R)2012 and A2:2010(R)2012, CSA CAN/CSA-C22.2  
NO 60601-1:14**

Additional Standards: **N/A**

Additional Information: **See the UL Online Certifications Directory at  
<https://iqulprospector.com> for additional information**

**This Certificate of Compliance does not provide authorization to apply the UL Recognized Component Mark. Only the UL Follow Up Services Procedure provides authorization to apply the UL Mark.**

**Only those products bearing the UL Recognized Component Mark should be considered as being UL Certified and covered under UL's Follow-Up Services.**

**Look for the UL Recognized Component Mark on the product.**



Bruce Mahrenholz, Director North American Certification Program

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at <http://ul.com/aboutul/locations/>