



Ref. Certif. No.

DK-130636-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

|   |   |
|---|---|
| Product   | Open frame medical power supply unit (PSU)  |
| Name and address of the applicant   | SELEC CONTROLS PVT LTD<br>EL 27/1, ELECTRONIC ZONE TTC<br>INDUSTRIAL AREA, MIDC<br>MAHAPE NAVI MUMBAI, Maharashtra 400709<br>INDIA  |
| Name and address of the manufacturer                                      | SELEC CONTROLS PVT LTD<br>EL 27/1, ELECTRONIC ZONE TTC<br>INDUSTRIAL AREA, MIDC<br>MAHAPE NAVI MUMBAI, Maharashtra 400709<br>INDIA  |
| Name and address of the factory   | SELEC CONTROLS PVT LTD<br>EL 30, ELECTRONIC ZONE , TTC INDUSTRIAL<br>AREA, MIDC MAHAPE,<br>NAVI MUMBIA, MAHARASTTRA- 400710<br>INDIA  |
| Note: When more than one factory, please report on page 2                 | <input type="checkbox"/> Additional Information on page 2   |
| Ratings and principal characteristics                                     | Input: 100-240 V ~, 47- 63 Hz, 1.4 A to 0.8 An<br>Output: see test report for details.  |
| Trademark (if any)  |   |
| Customer's Testing Facility (CTF) Stage used                              |   |
| Model / Type Ref.   | OPS2x4-60-XXX-VCYY-Z-CU<br><input checked="" type="checkbox"/> Additional Information on page 2   |
| Additional information (if necessary may also be reported on page 2)      | <b>Additionally evaluated to:</b> EN 60601-1:2006, EN 60601-1:2006/A1:2013, EN 60601-1:2006/A12:2014<br>National Differences specified in the CB Test Report.<br>Class II<br><br>The risk management requirements of the standard were not addressed<br><input type="checkbox"/> Additional Information on page 2 |
| A sample of the product was tested and found to be in conformity with     | IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012  |
| As shown in the Test Report Ref. No. which forms part of this Certificate | E525579-D1001-1/A0/C0-CB issued on 2022-07-28   |

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2022-08-03

Signature: Jan-Erik Storgaard



Ref. Certif. No.

**DK-130636-UL**

**Additional Model Detail(s):**

**OPS 60 Series:** OPS2x4-60-XXX-VCYY-Z-CU,

As per model Nomenclature - OPS2x4-60-XXX-VCYY-Z-CU

OPS- Represent Series name (Open frame power supply)

2x4 -Represent dimension in inches ( 2width x 4 Length).

60- Represent the maximum output power of family in watts.

XXX- Represent the typical output voltage in VDC .( First two XX represents the typical output voltage in VDC (05 VDC to 48V DC) and last X represent fractional value of output voltage, can be blank or any number from 1 to 9.).

V- Represent the input voltage range ( Blank for 100-240 V AC ).

C- Represent the connector type ( Any alphabet indicating type/ make of output connector).

YY- Represent the minor output voltage Variation ( Blank – Standard product or Any alphanumeric combination indicating minor output voltage variation in specification).

Z-Represent Classification of the device (2 for Class II (without earth)).

CU- Represent the Selec internal code ( UL, cUL approved).

**Additional information (if necessary)**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2022-08-03

Signature:

Jan-Erik Storgaard