



Ref. Certif. No.

DK-130595-A1-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

| | |
|---|---|
| Product | Open frame Medical power supply unit (PSU) |
| Name and address of the applicant | SELEC CONTROLS PVT LTD EL 27/1, ELECTRONIC ZONE TTC INDUSTRIAL AREA, MIDC MAHAPE NAVI MUMBAI, Maharashtra 400709 India |
| Name and address of the manufacturer | SELEC CONTROLS PVT LTD EL 27/1, ELECTRONIC ZONE TTC INDUSTRIAL AREA, MIDC MAHAPE NAVI MUMBAI, Maharashtra 400709 India |
| Name and address of the factory | Selec Controls Pvt Ltd EL30, Electronic Zone TTC Industrial Area MIDC Mahape Navi Mumbai, Maharashtra 400710 India |
| Note: When more than one factory, please report on page 2 | <input type="checkbox"/> Additional Information on page 2 |
| Ratings and principal characteristics | Input: 100-240 V ~, 47- 63 Hz, 0.85 A to 0.42 A Output: See CB Test Report for rating details. |
| Trademark (if any) | |
| Customer's Testing Facility (CTF) Stage used | |
| Model / Type Ref. | OPS2x3-40-XXX-VCYY-Z-CU, Class I <input checked="" type="checkbox"/> Additional Information on page 2 |
| Additional information (if necessary may also be reported on page 2) | Additionally evaluated to: EN 60601-1:2006, EN 60601-1:2006/A1:2013, EN 60601-1:2006/A12:2014 National Differences specified in the CB Test Report. The certificate was revised to include corrections. The risk management requirements of the standard were not addressed <input checked="" type="checkbox"/> Additional Information on page 2 |
| A sample of the product was tested and found to be in conformity with | IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 |
| As shown in the Test Report Ref. No. which forms part of this Certificate | E525579-D1002-1/A0/C0-CB issued on 2022-08-05 |

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2022-08-08
Original Issue Date: 2022-08-05

Signature: Jan-Erik Storgaard



Ref. Certif. No.

DK-130595-A1-UL

Additional Model Detail(s):

OPS 40 series: OPS2x3-40-XXX-VCYY-Z-CU, As per model Nomenclature - OPS2x4-40-XXX-VCYY-Z-CU

OPS- Represent Series name (Open frame power supply)

2x3 -Represent dimension in inches (2width x 3 Length).

40- Represent the maximum output power of family in watts.

XXX- Represent the typical output voltage in VDC .(First two XX represents the typical output voltage in VDC (05 V DC to 48 V DC) and last X represent fractional value of output voltage, can be blank or any number from 1 to 9.).

V- Represent the input voltage range (Blank for 100-240 V AC).

C- Represent the connector type (Any alphabet indicating type/ make of output connector).

YY- Represent the minor output voltage Variation (Blank – Standard product or Any alphanumeric combination indicating minor output voltage variation in specification).

Z-Represent Classification of the device (1 for Class 1 (with earth)).

CU- Represent the Selec internal code (UL, cUI approved).

Summary of Modifications:

This is correction to certificate only to add Class I information that was previously omitted.

Additional information (if necessary)



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2022-08-08

Original Issue Date: 2022-08-05

Signature:

Jan-Erik Storgaard